

BAPAT Proficiency Sample Evaluation Criteria

PT Sample Evaluation Criteria	Penalty Points
Failing to submit analysis results	Automatic "Fail"
Failing to report asbestos in a positive sample (FN)	100
Reporting 1% or greater asbestos in a blank sample (FP)	100
Failing to report the correct asbestos type, in a one-type sample (ID)	50
Failing to report second asbestos type, in a multi-type sample (ID)	50
Failing to report third asbestos type, in a multi-type sample (ID)	25
Reporting 1% or greater incorrect asbestos type (ID)	45/type
Reporting trace asbestos in a blank or trace incorrect asbestos type(s) in a positive sample	No Penalty
Failing to report trace asbestos when RTI QC confirms it in all samples	25/type
Failing to report asbestos when RTI QC does not confirm it in all samples	No penalty
Per sample, first asbestos quantitation outside acceptance range	20
Per sample, second asbestos quantitation outside acceptance range	10
Per sample, third asbestos quantitation outside acceptance range	5
Passing score for a round	Less than 100

Effective Date: January 30, 2007

BAPAT_PTSampleEvaluationCriteria_2007_01_30

Author: American Industrial Hygiene Association

Page 1 of 1